

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070220\
 Data File : VV017334.D
 Acq On : 03 Jul 2020 02:37
 Operator : SY/MD
 Sample : L3150-17
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jul 03 08:03:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 03 06:15:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	157170	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	153505	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	72599	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	41728	4.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.40%
7) Chloroethane-d5	1.58	69	33616	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.00%
11) 1,1-Dichloroethene-d2	2.12	63	67974	3.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.80%
20) 2-Butanone-d5	3.95	46	103531	43.06	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.12%
24) Chloroform-d	4.38	84	95636	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.06	65	54927	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.08	84	174053	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
36) 1,2-Dichloropropane-d6	6.10	67	49747	4.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.40%
41) Toluene-d8	7.34	98	158486	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.80%
45) trans-1,3-Dichloropropene-	7.65	79	20420	4.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.20%
48) 2-Hexanone-d5	8.11	63	66234	35.01	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.02%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	35121	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	62429	4.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.00%


Target Compounds					Ovalue
16) Methylene chloride	2.53	84	4788	0.593 ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070220\
Data File : VV017334.D
Acq On : 03 Jul 2020 02:37
Operator : SY/MD
Sample : L3150-17
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jul 03 08:03:35 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 03 06:15:26 2020
Response via : Initial Calibration





#16
 Methvlene chloride
 Concen: 0.593 ug/L
 RT: 2.53 min Scan# 447
 Delta R.T. 0.00 min
 Lab File: VV017334.D
 Acq: 03 Jul 2020 02:37

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Tot Ion: 84 Resp: 4788
 Ion Ratio Lower Upper
 84 100
 86 59.6 45.8 85.0
 49 118.8 89.6 166.4

