

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV070225\
 Data File : VV038839.D
 Acq On : 02 Jul 2025 14:50
 Operator : SY/MD
 Sample : VV0702WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0702WBL01

Quant Time: Jul 03 02:48:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 02:44:44 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.596	168	466070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	800338	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	747574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	357107	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	295406	56.554	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	113.100%
35) Dibromofluoromethane	4.500	113	262999	51.879	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	103.760%
50) Toluene-d8	7.236	98	763017	46.520	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	93.040%
62) 4-Bromofluorobenzene	10.001	95	284527	42.937	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	85.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV070225\
Data File : VV038839.D
Acq On : 02 Jul 2025 14:50
Operator : SY/MD
Sample : VV0702WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0702WBL01

Quant Time: Jul 03 02:48:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 02:44:44 2025
Response via : Initial Calibration

