

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV070225\
 Data File : VV038841.D
 Acq On : 02 Jul 2025 15:34
 Operator : SY/MD
 Sample : VV0702WBL03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0702WBL03

Quant Time: Jul 03 03:01:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 02:44:44 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	413719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	724343	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	672085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	322277	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	254805	54.954	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	109.900%
35) Dibromofluoromethane	4.506	113	224163	48.857	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	97.720%
50) Toluene-d8	7.239	98	632653	42.619	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	85.240%#
62) 4-Bromofluorobenzene	10.001	95	235797	39.316	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	78.640%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV070225\
Data File : VV038841.D
Acq On : 02 Jul 2025 15:34
Operator : SY/MD
Sample : VV0702WBL03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0702WBL03

Quant Time: Jul 03 03:01:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 02:44:44 2025
Response via : Initial Calibration

