

Data File : VV011823.D
Acq On : 03 Jul 2019 10:54
Operator : SY/MD
Sample : K3608-07DL 4X
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAML6DL

Quant Time: Jul 04 00:07:05 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 03 23:52:56 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	163675	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	156136	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	67952	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	50304	4.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.40%
7) Chloroethane-d5	1.58	69	42728	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.00%
11) 1,1-Dichloroethene-d2	2.13	63	75860	3.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.40%
20) 2-Butanone-d5	3.95	46	131795	48.45	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.90%
24) Chloroform-d	4.40	84	99627	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
26) 1,2-Dichloroethane-d4	5.08	65	54584	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.09	84	196753	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.12	67	62894	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.36	98	169883	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
43) trans-1,3-Dichloropropene-	7.66	79	20196	4.52	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.40%
46) 2-Hexanone-d5	8.13	63	99350	48.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.48%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	41365	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	62007	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds					Qvalue
34) Trichloroethene	5.96	95	86844	7.181 ug/L	99
47) Tetrachloroethene	8.02	164	14187	1.491 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011823.D
Acq On : 03 Jul 2019 10:54
Operator : SY/MD
Sample : K3608-07DL 4X
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAML6DL

Quant Time: Jul 04 00:07:05 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 03 23:52:56 2019
Response via : Initial Calibration

