

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW070522\
 Data File : VW026639.D
 Acq On : 05 Jul 2022 13:09
 Operator : SY/MD
 Sample : N3577-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB323

Quant Time: Jul 06 00:23:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 06 00:22:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	229852	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	224783	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	93058	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	52554	2.843	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	56.800%	
7) Chloroethane-d5	1.581	69	50102	3.374	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	67.400%	
11) 1,1-Dichloroethene-d2	2.117	63	74817	2.278	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	45.600%#	
20) 2-Butanone-d5	3.915	46	99823	54.317	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.640%	
24) Chloroform-d	4.359	84	119852	3.966	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	79.400%	
26) 1,2-Dichloroethane-d4	5.040	65	49583	4.260	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.200%	
32) Benzene-d6	5.056	84	207930	3.220	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	64.400%#	
36) 1,2-Dichloropropane-d6	6.072	67	68614	3.865	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	77.200%	
41) Toluene-d8	7.317	98	161166	2.776	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	55.600%#	
43) trans-1,3-Dichloroprop...	7.629	79	16720	3.643	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	72.800%	
46) 2-Hexanone-d5	8.092	63	72319	41.371	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	82.740%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	50714	4.366	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	87.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	59841	3.886	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	77.800%#	
Target Compounds						
3) Chloromethane	1.253	50	4488	0.234	ug/L	93
16) Methylene chloride	2.516	84	3174	0.193	ug/L	80
22) cis-1,2-Dichloroethene	3.928	96	13486	0.857	ug/L	98
34) Trichloroethene	5.918	95	69216	4.011	ug/L	99
47) Tetrachloroethene	7.976	164	3618	0.263	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070522\
 Data File : VV026639.D
 Acq On : 05 Jul 2022 13:09
 Operator : SY/MD
 Sample : N3577-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB323

Quant Time: Jul 06 00:23:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 06 00:22:14 2022
 Response via : Initial Calibration

