

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070620\
 Data File : VV017353.D
 Acq On : 06 Jul 2020 15:41
 Operator : SY/MD
 Sample : L3168-01DL 4X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AH6DL

Quant Time: Jul 07 03:10:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 07 01:14:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	164686	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	158492	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	77438	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	41575	4.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.20%
7) Chloroethane-d5	1.58	69	35236	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.00%
11) 1,1-Dichloroethene-d2	2.13	63	68320	3.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.60%
20) 2-Butanone-d5	3.94	46	96898	38.47	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.94%
24) Chloroform-d	4.38	84	94047	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
26) 1,2-Dichloroethane-d4	5.06	65	53009	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
32) Benzene-d6	5.08	84	175291	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.20%
36) 1,2-Dichloropropane-d6	6.10	67	49442	4.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.40%
41) Toluene-d8	7.34	98	158895	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
45) trans-1,3-Dichloropropene-	7.65	79	20551	4.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.20%
48) 2-Hexanone-d5	8.11	63	66067	33.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	67.64%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	34402	4.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	63827	4.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.20%

Target Compounds					Ovalue
49) Tetrachloroethene	8.00	164	56909	5.817 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070620\
Data File : VV017353.D
Acq On : 06 Jul 2020 15:41
Operator : SY/MD
Sample : L3168-01DL 4X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AH6DL

Quant Time: Jul 07 03:10:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 07 01:14:17 2020
Response via : Initial Calibration

