

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070623\
 Data File : VV031652.D
 Acq On : 06 Jul 2023 14:03
 Operator : SY/MD
 Sample : VIBLK221
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK221

Quant Time: Jul 07 02:37:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 07 02:34:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	168481	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	171504	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	89921	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	44580	39.758	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.520%
7) Chloroethane-d5	1.532	69	42789	43.735	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.460%
11) 1,1-Dichloroethene-d2	2.060	65	19474	33.698	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.400%
21) 2-Butanone-d5	3.786	46	115259	126.945	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	126.940%
24) Chloroform-d	4.252	84	127939	45.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.120%
26) 1,2-Dichloroethane-d4	4.947	65	89274	50.811	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.620%
32) Benzene-d6	4.966	84	192329	39.841	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.680%
36) 1,2-Dichloropropane-d6	5.992	67	78925	46.128	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.260%
41) Toluene-d8	7.249	98	149324	33.708	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	67.420%#
43) trans-1,3-Dichloroprop...	7.558	79	32906	39.451	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.900%
47) 2-Hexanone-d5	8.030	63	75451	111.687	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.690%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	127168	49.240	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.480%
66) 1,2-Dichlorobenzene-d4	11.567	152	81209	47.378	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070623\
Data File : VV031652.D
Acq On : 06 Jul 2023 14:03
Operator : SY/MD
Sample : VIBLK221
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK221

Quant Time: Jul 07 02:37:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 07 02:34:55 2023
Response via : Initial Calibration

