

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW070722\
 Data File : VW026679.D
 Acq On : 07 Jul 2022 16:07
 Operator : SY/MD
 Sample : N3601-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUA8

Quant Time: Jul 08 03:02:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 08 02:58:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	192485	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	182138	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	79288	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	67771	4.378	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	87.600%	
7) Chloroethane-d5	1.580	69	50914	4.094	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	81.800%	
11) 1,1-Dichloroethene-d2	2.124	63	138377	5.031	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	100.600%	
20) 2-Butanone-d5	3.915	46	89491	58.148	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	116.300%	
24) Chloroform-d	4.362	84	103203	4.078	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.600%	
26) 1,2-Dichloroethane-d4	5.047	65	42829	4.394	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.800%	
32) Benzene-d6	5.059	84	204909	3.917	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	78.400%	
36) 1,2-Dichloropropane-d6	6.079	67	62280	4.329	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	86.600%	
41) Toluene-d8	7.320	98	158639	3.372	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	67.400%#	
43) trans-1,3-Dichloroprop...	7.632	79	13915	3.742	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.800%	
46) 2-Hexanone-d5	8.095	63	70805	49.989	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.980%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39267	4.172	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	83.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	54286	4.137	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	82.800%	
Target Compounds						
5) Vinyl chloride	1.323	62	1420643	83.239	ug/L	93
12) 1,1-Dichloroethene	2.130	96	112784	8.675	ug/L	84
18) trans-1,2-Dichloroethene	2.770	96	3362813	253.534	ug/L	96
22) cis-1,2-Dichloroethene	3.921	96	8535949	648.088	ug/L	96
34) Trichloroethene	5.924	95	13002198	929.991	ug/L	97
47) Tetrachloroethene	7.982	164	7910074	710.622	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070722\
Data File : VV026679.D
Acq On : 07 Jul 2022 16:07
Operator : SY/MD
Sample : N3601-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUA8

Quant Time: Jul 08 03:02:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 08 02:58:15 2022
Response via : Initial Calibration

