

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070722\
 Data File : VV026682.D
 Acq On : 07 Jul 2022 17:17
 Operator : SY/MD
 Sample : N3601-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUB4

Quant Time: Jul 08 03:03:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 08 02:58:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	175150	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	168157	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68327	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	55907	3.969	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.400%
7) Chloroethane-d5	1.581	69	52386	4.629	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.600%
11) 1,1-Dichloroethene-d2	2.118	63	83861	3.351	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.000%
20) 2-Butanone-d5	3.928	46	79512	56.778	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.560%
24) Chloroform-d	4.362	84	111314	4.834	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.600%
26) 1,2-Dichloroethane-d4	5.044	65	44035	4.965	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%
32) Benzene-d6	5.060	84	201026	4.162	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.200%
36) 1,2-Dichloropropane-d6	6.076	67	63128	4.753	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.320	98	161533	3.719	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.400%
43) trans-1,3-Dichloroprop...	7.629	79	12625	3.677	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.600%
46) 2-Hexanone-d5	8.095	63	58884	45.029	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.060%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	41593	4.787	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	54303	4.802	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%
Target Compounds						
14) Carbon disulfide	2.307	76	9408	0.301	ug/L	96
18) trans-1,2-Dichloroethene	2.773	96	3073	0.255	ug/L	94
22) cis-1,2-Dichloroethene	3.931	96	9463	0.790	ug/L	85
34) Trichloroethene	5.925	95	28585	2.215	ug/L	96
47) Tetrachloroethene	7.979	164	8576	0.835	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070722\
Data File : VV026682.D
Acq On : 07 Jul 2022 17:17
Operator : SY/MD
Sample : N3601-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUB4

Quant Time: Jul 08 03:03:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 08 02:58:15 2022
Response via : Initial Calibration

