

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV070725\
 Data File : VV038866.D
 Acq On : 07 Jul 2025 12:28
 Operator : SY/MD
 Sample : VV0707WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0707WBL02

Quant Time: Jul 08 02:38:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V070225W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 08 02:37:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	426999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	738137	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	687689	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	341949	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	279653	58.437	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	116.880%
35) Dibromofluoromethane	4.506	113	243737	52.131	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	104.260%
50) Toluene-d8	7.239	98	701922	46.401	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	92.800%
62) 4-Bromofluorobenzene	10.001	95	298109	48.777	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	97.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV070725\
Data File : VV038866.D
Acq On : 07 Jul 2025 12:28
Operator : SY/MD
Sample : VV0707WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0707WBL02

Quant Time: Jul 08 02:38:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V070225W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 08 02:37:09 2025
Response via : Initial Calibration

