

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV070725\
 Data File : VV038867.D
 Acq On : 07 Jul 2025 12:50
 Operator : SY/MD
 Sample : VV0707WBL03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0707WBL03

Quant Time: Jul 08 02:39:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V070225W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 08 02:37:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	408431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	716173	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	681665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	322968	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	269615	58.901	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	117.800%
35) Dibromofluoromethane	4.506	113	241138	53.156	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	106.320%
50) Toluene-d8	7.239	98	677242	46.143	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	92.280%
62) 4-Bromofluorobenzene	10.001	95	287126	48.421	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	96.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV070725\
Data File : VV038867.D
Acq On : 07 Jul 2025 12:50
Operator : SY/MD
Sample : VV0707WBL03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0707WBL03

Quant Time: Jul 08 02:39:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V070225W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 08 02:37:09 2025
Response via : Initial Calibration

