

Data File : VV011836.D
Acq On : 08 Jul 2019 16:11
Operator : SY/MD
Sample : VSTD2.006
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002061

Quant Time: Jul 09 01:46:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 09 01:42:50 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	8364	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	7777	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3609	0.50	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Vinyl Chloride-d3	1.32	65	18330	2.19	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.07	65	14460	2.24	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.12	67	15621	2.16	ug/L	0.00
8) Toluene-d8	7.35	98	42193	2.29	ug/L	0.00
10) 1,1,2,2-Tetrachloroethane-	10.25	84	9723	2.05	ug/L	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Vinyl chloride	1.32	62	18380	2.305	ug/L	98
6) Trichloroethene	5.95	95	14806	2.199	ug/L	98
9) 1,2-Dibromoethane	8.40	107	8211	2.165	ug/L #	98
12) 1,2,3-Trichloropropane	10.30	75	8139	1.934	ug/L	99
13) 1,2-Dibromo-3-chloropropan	12.46	75	1332	1.722	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011836.D
Acq On : 08 Jul 2019 16:11
Operator : SY/MD
Sample : VSTD2.006
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002061

Quant Time: Jul 09 01:46:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 09 01:42:50 2019
Response via : Initial Calibration





#3

Vinyl chloride

Concen: 2.305 ug/L

RT: 1.32 min Scan# 60

Delta R.T. 0.00 min

Lab File: VV011836.D

Acq: 08 Jul 2019 16:11

Instrument :

MSVOA_V

ClientSampled :





VSTD002061

Tgt Ion: 62 Resp: 18380

Ion Ratio Lower Upper

62 100

64 32.3 23.2 43.2



#6

Trichloroethene

Concen: 2.199 ug/L

RT: 5.95 min Scan# 419

Delta R.T. 0.00 min

Lab File: VV011836.D

Acq: 08 Jul 2019 16:11

Tgt Ion: 95 Resp: 14806




Ion Ratio Lower Upper


95 100

97 63.8 46.2 85.8

132 85.5 59.0 109.6

130 92.7 64.3 119.5





#9

1,2-Dibromoethane

Concen: 2.165 ug/L

RT: 8.40 min Scan# 548

Delta R.T. 0.00 min

Lab File: VV011836.D

Acq: 08 Jul 2019 16:11

Instrument :

MSVOA_V

ClientSampleId :

VSTD002061


Tgt Ion: 107 Resp: 8211

Ion Ratio Lower Upper

107 100

109 96.5 67.7 125.7




188 6.2 9.8 14.8#



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V

Ion 188.00 (187.70 to 188.70): V



#12

1,2,3-Trichloropropane

Concen: 1.934 ug/L

RT: 10.30 min Scan# 661

Delta R.T. 0.00 min

Lab File: VV011836.D

Acq: 08 Jul 2019 16:11


Tgt Ion: 75 Resp: 8139

Ion Ratio Lower Upper

75 100

110 39.3 32.2 48.4



77 32.2 25.3 37.9




Abundance Ion 75.00 (74.70 to 75.70): VV0

Ion 110.00 (109.70 to 110.70): V

Ion 77.00 (76.70 to 77.70): VV0





#13

1,2-Dibromo-3-chloropropane

Concen: 1.722 ug/L

RT: 12.46 min Scan# 787

Delta R.T. 0.00 min

Lab File: VV011836.D

Acq: 08 Jul 2019 16:11

Instrument :

MSVOA_V

ClientSampleId :

VSTD002061


Tgt Ion: 75 Resp: 1332

Ion Ratio Lower Upper

75 100

157 76.7 62.2 93.4

155 65.4 55.6 83.4



Abundance Ion 75.00 (74.70 to 75.70): VV011836.D (-769) (-)

Ion 157.00 (156.70 to 157.70): VV011836.D (-769) (-)

Ion 155.00 (154.70 to 155.70): VV011836.D (-769) (-)

