

Data File : VV011834.D
Acq On : 08 Jul 2019 13:35
Operator : SY/MD
Sample : VSTD0.504
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5041

Quant Time: Jul 09 01:46:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 09 01:42:50 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	9296	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	8362	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3859	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	4621	0.50	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.07	65	3775	0.53	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.12	67	3971	0.51	ug/L	0.00
8) Toluene-d8	7.35	98	10333	0.52	ug/L	0.00
10) 1,1,2,2-Tetrachloroethane-	10.25	84	2482	0.49	ug/L	0.00

Target Compounds

						Qvalue
3) Vinyl chloride	1.32	62	4361	0.492	ug/L	100
6) Trichloroethene	5.95	95	3445	0.476	ug/L	100
9) 1,2-Dibromoethane	8.40	107	1938	0.475	ug/L	100
12) 1,2,3-Trichloropropane	10.30	75	1979	0.440	ug/L	100
13) 1,2-Dibromo-3-chloropropan	12.46	75	338	0.409	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011834.D
Acq On : 08 Jul 2019 13:35
Operator : SY/MD
Sample : VSTD0.504
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5041

Quant Time: Jul 09 01:46:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 09 01:42:50 2019
Response via : Initial Calibration

