

Data File : VV011837.D
Acq On : 08 Jul 2019 16:39
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV011

Quant Time: Jul 09 02:06:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 09 02:02:54 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	9034	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	8203	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3809	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	4597	0.47	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	94.00%
4) 1,2-Dichloroethane-d4	5.07	65	3817	0.52	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.00%
7) 1,2-Dichloropropane-d6	6.12	67	4130	0.52	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.00%
8) Toluene-d8	7.35	98	10542	0.51	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.00%
10) 1,1,2,2-Tetrachloroethane-	10.25	84	2561	0.52	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	104.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.32	62	4463	0.476 ug/L	99
6) Trichloroethene	5.95	95	3563	0.472 ug/L	100
9) 1,2-Dibromoethane	8.40	107	2006	0.490 ug/L	100
12) 1,2,3-Trichloropropane	10.30	75	2051	0.494 ug/L	99
13) 1,2-Dibromo-3-chloropropan	12.46	75	343	0.450 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011837.D
Acq On : 08 Jul 2019 16:39
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV011

Quant Time: Jul 09 02:06:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 09 02:02:54 2019
Response via : Initial Calibration

