

Method Path : Z:\VOASRV\HPCHEM1\MSV0A_V\METHOD\
 Method File : SOMVTR070819SIM.M
 Title : TRACE VOA SOM01.0
 Last Update : Tue Jul 09 01:52:46 2019
 Response Via : Initial Calibration

Calibration Files

0.5 =VV011834.D 0.05=VV011832.D 0.1 =VV011833.D
 1.0 =VV011835.D 2 =VV011836.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.497	0.623	0.509	0.550	0.548	0.546	9.06
3) T	Vinyl chloride	0.469	0.545	0.502	0.530	0.549	0.519	6.47
4) S	1,2-Dichloroethane-	0.406	0.399	0.380	0.429	0.432	0.409	5.31
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.412	0.491	0.463	0.459	0.476	0.460	6.45
7) S	1,2-Dichloropropane	0.475	0.484	0.456	0.500	0.502	0.483	3.96
8) S	Toluene-d8	1.236	1.250	1.164	1.312	1.356	1.264	5.86
9) T	1,2-Dibromoethane	0.232	0.256	0.240	0.255	0.264	0.249	5.22
10) S	1,1,2,2-Tetrachloro	0.297	0.302	0.292	0.311	0.313	0.303	2.93
11) I	1,4-Dichlorobenzene-d	-----ISTD-----						
12)	1,2,3-Trichloroprop	0.513	0.538	0.547	0.560	0.564	0.544	3.75
13) T	1,2-Dibromo-3-chlor	0.088	0.121	0.105	0.094	0.092	0.100	13.50

(#) = Out of Range