

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW070822\
 Data File : VW026695.D
 Acq On : 08 Jul 2022 14:21
 Operator : SY/MD
 Sample : N3601-06DL 200X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUB1DL

Quant Time: Jul 08 23:56:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 08 23:50:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	176350	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	174520	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	66219	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	57040	4.022	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	80.400%	
7) Chloroethane-d5	1.581	69	54080	4.747	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.000%	
11) 1,1-Dichloroethene-d2	2.118	63	82177	3.261	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	65.200%	
20) 2-Butanone-d5	3.931	46	65289	46.304	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	92.600%	
24) Chloroform-d	4.362	84	118223	5.099	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.000%	
26) 1,2-Dichloroethane-d4	5.043	65	48039	5.380	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.600%	
32) Benzene-d6	5.060	84	210762	4.204	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.000%	
36) 1,2-Dichloropropane-d6	6.076	67	69858	5.068	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.400%	
41) Toluene-d8	7.320	98	155243	3.444	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	68.800%#	
43) trans-1,3-Dichloroprop...	7.632	79	12491	3.506	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	70.200%	
46) 2-Hexanone-d5	8.095	63	64682	47.660	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.320%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	49125	5.448	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	109.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	57532	5.250	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.000%	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.519	84	9943	0.786	ug/L	96
18) trans-1,2-Dichloroethene	2.777	96	10931	0.900	ug/L	93
22) cis-1,2-Dichloroethene	3.928	96	71665	5.939	ug/L	99
34) Trichloroethene	5.918	95	388836	29.026	ug/L	97
47) Tetrachloroethene	7.976	164	67931	6.369	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070822\
Data File : VV026695.D
Acq On : 08 Jul 2022 14:21
Operator : SY/MD
Sample : N3601-06DL 200X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUB1DL

Quant Time: Jul 08 23:56:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 08 23:50:05 2022
Response via : Initial Calibration

