

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW070822\
 Data File : VW026705.D
 Acq On : 08 Jul 2022 18:48
 Operator : SY/MD
 Sample : N3601-21 50X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUB6

Quant Time: Jul 08 23:59:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 08 23:50:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	178846	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	177211	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68911	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	53859	3.745	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	74.800%	
7) Chloroethane-d5	1.580	69	50734	4.391	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	87.800%	
11) 1,1-Dichloroethene-d2	2.117	63	75500	2.954	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	59.000%#	
20) 2-Butanone-d5	3.931	46	66356	46.404	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	92.800%	
24) Chloroform-d	4.362	84	114254	4.859	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.200%	
26) 1,2-Dichloroethane-d4	5.043	65	45711	5.048	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.000%	
32) Benzene-d6	5.059	84	198137	3.893	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	77.800%	
36) 1,2-Dichloropropane-d6	6.075	67	61910	4.423	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	88.400%	
41) Toluene-d8	7.320	98	144759	3.162	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	63.200%#	
43) trans-1,3-Dichloroprop...	7.632	79	12824	3.544	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	70.800%	
46) 2-Hexanone-d5	8.095	63	59706	43.325	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	86.660%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	46442	5.072	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.400%	
66) 1,2-Dichlorobenzene-d4	11.625	152	53853	4.722	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.400%	
Target Compounds						Qvalue
16) Methylene chloride	2.523	84	7543	0.588	ug/L	83
47) Tetrachloroethene	7.982	164	1525	0.141	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070822\
Data File : VV026705.D
Acq On : 08 Jul 2022 18:48
Operator : SY/MD
Sample : N3601-21 50X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUB6

Quant Time: Jul 08 23:59:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 08 23:50:05 2022
Response via : Initial Calibration

