

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW070822\
 Data File : VW026690.D
 Acq On : 08 Jul 2022 12:15
 Operator : SY/MD
 Sample : N3601-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUC0

Quant Time: Jul 08 23:54:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 08 23:50:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	234413	5.000	ug/L	-0.02
28) Chlorobenzene-d5	8.857	117	161395	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62564	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	61547	3.265	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.200%
7) Chloroethane-d5	1.574	69	56090	3.704	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.000%
11) 1,1-Dichloroethene-d2	2.111	63	88401	2.639	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.800%#
20) 2-Butanone-d5	3.954	46	119752	63.893	ug/L	0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.780%
24) Chloroform-d	4.365	84	127891	4.150	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
26) 1,2-Dichloroethane-d4	5.040	65	56807	4.786	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%
32) Benzene-d6	5.047	84	261318	5.637	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.800%
36) 1,2-Dichloropropane-d6	6.053	67	69377	5.442	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.800%
41) Toluene-d8	7.310	98	167651	4.022	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
43) trans-1,3-Dichloroprop...	7.632	79	10664	3.236	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.800%
46) 2-Hexanone-d5	8.117	63	51121	40.731	ug/L	0.03
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.460%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	41751	5.006	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	49825	4.812	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070822\
Data File : VV026690.D
Acq On : 08 Jul 2022 12:15
Operator : SY/MD
Sample : N3601-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUC0

Quant Time: Jul 08 23:54:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 08 23:50:05 2022
Response via : Initial Calibration

