

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070918\
 Data File : VV006562.D
 Acq On : 09 Jul 2018 16:48
 Operator : SY/MD
 Sample : J3792-11RE
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB0H6RE

Quant Time: Jul 09 23:05:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jul 09 22:52:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	344445	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	315180	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	149820	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80164	4.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.60%
7) Chloroethane-d5	1.58	69	74315	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.00%
11) 1,1-Dichloroethene-d2	2.13	63	106657	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	3.97	46	207713	43.67	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.34%
24) Chloroform-d	4.41	84	162978	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.09	65	77967	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
32) Benzene-d6	5.10	84	339119	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.12	67	108932	4.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.40%
41) Toluene-d8	7.37	98	295528	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
43) trans-1,3-Dichloropropene-	7.67	79	35775	4.52	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.40%
46) 2-Hexanone-d5	8.14	63	179850	45.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.06%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	81966	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	122510	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds

					Ovalue
13) Acetone	2.22	43	4855	1.61 ug/L	96
33) Benzene	5.15	78	45709	0.49 ug/L	100
51) Chlorobenzene	8.93	112	210333	3.29 ug/L	99
63) 1,4-Dichlorobenzene	11.32	146	267277	5.56 ug/L	99
65) 1,2-Dichlorobenzene	11.70	146	27276	0.60 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070918\
Data File : VV006562.D
Acq On : 09 Jul 2018 16:48
Operator : SY/MD
Sample : J3792-11RE
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB0H6RE

Quant Time: Jul 09 23:05:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jul 09 22:52:57 2018
Response via : Initial Calibration

