

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070919\
 Data File : VV011845.D
 Acq On : 09 Jul 2019 12:14
 Operator : SY/MD
 Sample : MDL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL01

Quant Time: Jul 10 01:47:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 10 01:45:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	8425	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	7703	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3640	0.50	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.32	65	4131	0.45	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	90.00%
4) 1,2-Dichloroethane-d4	5.07	65	2893	0.42	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	84.00%
7) 1,2-Dichloropropane-d6	6.12	67	3372	0.45	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	90.00%
8) Toluene-d8	7.35	98	9013	0.46	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.00%
10) 1,1,2,2-Tetrachloroethane-	10.25	84	1885	0.40	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	80.00%

Target Compounds					Qvalue
3) Vinyl chloride	1.32	62	290	0.033 ug/L #	37
6) Trichloroethene	5.95	95	232	0.033 ug/L	88
9) 1,2-Dibromoethane	8.40	107	111	0.029 ug/L #	85
12) 1,2,3-Trichloropropane	10.30	75	107	0.027 ug/L #	91
13) 1,2-Dibromo-3-chloropropan	12.46	75	22	0.030 ug/L #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070919\
Data File : VV011845.D
Acq On : 09 Jul 2019 12:14
Operator : SY/MD
Sample : MDL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL01

Quant Time: Jul 10 01:47:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 10 01:45:06 2019
Response via : Initial Calibration

