

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070919\
 Data File : VV011846.D
 Acq On : 09 Jul 2019 12:50
 Operator : SY/MD
 Sample : MDL02
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL02

Quant Time: Jul 10 01:48:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 10 01:45:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	8629	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	7820	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3567	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	4724	0.50	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	100.00%
4) 1,2-Dichloroethane-d4	5.07	65	3626	0.51	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.00%
7) 1,2-Dichloropropane-d6	6.12	67	3927	0.52	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.00%
8) Toluene-d8	7.35	98	9621	0.49	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.00%
10) 1,1,2,2-Tetrachloroethane-	10.25	84	2257	0.48	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	96.00%

Target Compounds

					Qvalue	
3) Vinyl chloride	1.32	62	353	0.039	ug/L	# 36
6) Trichloroethene	5.95	95	251	0.035	ug/L	90
9) 1,2-Dibromoethane	8.40	107	132	0.034	ug/L	# 86
12) 1,2,3-Trichloropropane	10.30	75	132	0.034	ug/L	# 78
13) 1,2-Dibromo-3-chloropropan	12.46	75	26	0.036	ug/L	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070919\
Data File : VV011846.D
Acq On : 09 Jul 2019 12:50
Operator : SY/MD
Sample : MDL02
Misc : 25ML/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL02

Quant Time: Jul 10 01:48:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 10 01:45:06 2019
Response via : Initial Calibration

