

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070920\
 Data File : VV017396.D
 Acq On : 09 Jul 2020 14:29
 Operator : SY/MD
 Sample : L3254-10
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4A57

Quant Time: Jul 10 04:18:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM070820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 10 04:00:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	250185	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	239088	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	115939	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	60577	46.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.02%
7) Chloroethane-d5	1.58	69	45152	49.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.70%
11) 1,1-Dichloroethene-d2	2.12	63	101543	34.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.86%
21) 2-Butanone-d5	3.91	46	88212	99.16	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.16%
24) Chloroform-d	4.38	84	143738	47.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.42%
26) 1,2-Dichloroethane-d4	5.06	65	105352	50.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.90%
32) Benzene-d6	5.08	84	255716	49.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.02%
36) 1,2-Dichloropropane-d6	6.09	67	75182	50.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.10%
41) Toluene-d8	7.34	98	240386	48.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.82%
43) trans-1,3-Dichloropropene-	7.64	79	41632	48.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.18%
47) 2-Hexanone-d5	8.11	63	57248	93.79	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.79%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	96981	46.44	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.88%
64) 1,2-Dichlorobenzene-d4	11.65	152	102666	52.96	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.92%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070920\
Data File : VV017396.D
Acq On : 09 Jul 2020 14:29
Operator : SY/MD
Sample : L3254-10
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4A57

Quant Time: Jul 10 04:18:08 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM070820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 10 04:00:27 2020
Response via : Initial Calibration

