

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV070920\
 Data File : VV017415.D
 Acq On : 09 Jul 2020 22:14
 Operator : SY/MD
 Sample : L3253-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4A22

Quant Time: Jul 10 05:25:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM070820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 10 05:13:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	278513	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	265867	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	127539	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	58560	40.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.78%
7) Chloroethane-d5	1.58	69	45715	44.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.76%
11) 1,1-Dichloroethene-d2	2.12	63	98121	29.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.78%#
21) 2-Butanone-d5	3.91	46	87346	88.20	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.20%
24) Chloroform-d	4.38	84	140503	41.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.78%
26) 1,2-Dichloroethane-d4	5.06	65	103089	44.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.56%
32) Benzene-d6	5.08	84	247452	43.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.18%
36) 1,2-Dichloropropane-d6	6.09	67	72575	43.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.76%
41) Toluene-d8	7.34	98	231417	41.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.82%
43) trans-1,3-Dichloropropene-	7.64	79	40205	41.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.54%
47) 2-Hexanone-d5	8.11	63	58956	86.86	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.86%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	98099	42.24	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.48%
64) 1,2-Dichlorobenzene-d4	11.65	152	97357	45.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.30%


Target Compounds					Ovalue
13) Acetone	2.20	43	5896	4.374 ug/L	73


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV070920\
Data File : VV017415.D
Acq On : 09 Jul 2020 22:14
Operator : SY/MD
Sample : L3253-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4A22

Quant Time: Jul 10 05:25:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM070820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 10 05:13:25 2020
Response via : Initial Calibration





#13

Acetone

Concen: 4.374 ug/L

RT: 2.20 min Scan# 345

Delta R.T. 0.01 min

Lab File: VV017415.D

Acq: 09 Jul 2020 22:14

Instrument :
MSVOA_V
ClientSampleId :
F4A22

Tot Ion: 43 Resp: 5896

Ion Ratio Lower Upper

43 100

58 13.7 0.0 56.0

