

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071018\
 Data File : VV006579.D
 Acq On : 10 Jul 2018 19:56
 Operator : SY/MD
 Sample : J3919-07MSD
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AJ2MSD

Quant Time: Jul 11 04:02:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 11 03:34:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	315884	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	288041	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	126431	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	68232	3.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.40%
7) Chloroethane-d5	1.58	69	62921	4.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.00%
11) 1,1-Dichloroethene-d2	2.13	63	133738	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.80%
20) 2-Butanone-d5	3.95	46	197258	45.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.44%
24) Chloroform-d	4.41	84	141346	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.60%
26) 1,2-Dichloroethane-d4	5.09	65	71159	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
32) Benzene-d6	5.10	84	288514	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
36) 1,2-Dichloropropane-d6	6.12	67	94030	4.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.20%
41) Toluene-d8	7.36	98	247769	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
43) trans-1,3-Dichloropropene-	7.67	79	27713	3.83	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.60%
46) 2-Hexanone-d5	8.14	63	182330	49.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.90%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	73611	4.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	102274	4.70	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.00%

Target Compounds					Ovalue
12) 1,1-Dichloroethene	2.14	96	88330	4.47	ug/L 97
13) Acetone	2.19	43	3150	1.14	ug/L 89
33) Benzene	5.15	78	392904	4.57	ug/L 100
34) Trichloroethene	5.96	95	96554	4.38	ug/L 99
42) Toluene	7.44	91	409117	4.58	ug/L 98
51) Chlorobenzene	8.93	112	272661	4.66	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV071018\
Data File : VV006579.D
Acq On : 10 Jul 2018 19:56
Operator : SY/MD
Sample : J3919-07MSD
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AJ2MSD

Quant Time: Jul 11 04:02:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 11 03:34:14 2018
Response via : Initial Calibration

