


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071018\
Data File : VV006565.D
Acq On : 10 Jul 2018 11:38
Operator : SY/MD
Sample : BFB63
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB63

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
Title : TRACE VOA SOM01.0
Last Update : Wed Jul 11 03:34:14 2018



AutoFind: Scans 2871, 2872, 2873; Background Corrected with Scan 2859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.1	13122	PASS
75	95	30	80	45.5	39522	PASS
95	95	100	100	100.0	86909	PASS
96	95	5	9	6.7	5854	PASS
173	174	0.00	2	0.8	596	PASS
174	95	50	120	85.4	74189	PASS
175	174	5	9	7.0	5190	PASS
176	174	95	101	96.3	71474	PASS
177	176	5	9	6.4	4604	PASS

Scan 2859 (10.083 min): VV006565.D
BFB63

m/z	Abundance
39.80	159.0
43.80	763.0

Instrument :
MSVOA_V
ClientSampleId :
BFB63

m/z	Abundance
35.90	490.0
36.90	2803.0
37.90	2628.0
38.90	1163.0
39.90	285.0
43.90	980.0
44.90	508.0
46.90	949.0
47.80	425.0
48.90	2507.0
49.90	12793.0
50.90	4283.0
52.00	157.0
54.80	204.0
55.90	964.0
56.90	1691.0
59.80	632.0
60.90	3281.0
61.90	3325.0
62.90	2559.0
63.90	295.0
64.90	524.0
66.90	227.0
67.90	6868.0
68.90	8001.0
69.90	543.0
71.90	439.0
72.90	3224.0
73.90	12254.0
74.90	38112.0
76.00	3224.0
76.80	628.0
77.80	342.0
78.80	1847.0
79.90	487.0
80.80	1867.0
81.80	460.0
85.70	130.0
86.90	3848.0
87.90	3448.0
90.70	285.0
91.90	2035.0
92.90	3050.0
93.90	9292.0
94.90	84128.0
95.90	5804.0
97.00	140.0
103.80	253.0
105.80	281.0
106.90	110.0
115.80	254.0
116.80	449.0
117.80	267.0
118.80	402.0
127.70	251.0
128.80	118.0
129.70	209.0
134.80	113.0
136.80	123.0
140.90	631.0
142.80	747.0
144.80	117.0
145.70	129.0
148.00	185.0
154.80	185.0
156.80	134.0
171.80	620.0
172.90	606.0
173.80	69872.0
174.90	5035.0
175.80	67024.0
176.90	4486.0
177.90	133.0

Instrument :
MSVOA_V
ClientSampleId :
BFB63

m/z	Abundance
35.90	524.0
36.90	2926.0
37.90	2735.0
38.90	1190.0
39.90	238.0
43.90	1019.0
44.90	527.0
46.90	1001.0
47.80	435.0
48.90	2796.0
49.90	13505.0
50.90	4520.0
52.00	181.0
54.90	225.0
55.90	967.0
56.90	1833.0
59.90	652.0
60.90	3447.0
61.90	3568.0
62.90	2772.0
64.00	296.0
64.90	586.0
66.90	209.0
67.90	7463.0
68.90	8273.0
69.90	586.0
72.00	448.0
72.90	3443.0
73.90	13063.0
74.90	40920.0
76.00	3578.0
76.80	669.0
77.80	399.0
78.80	1936.0
79.90	529.0
80.80	2037.0
81.80	466.0
85.80	148.0
86.90	4150.0
87.90	3673.0
90.80	275.0
91.90	2370.0
92.90	3157.0
93.90	9715.0
94.90	89832.0
95.90	6081.0
97.00	143.0
103.80	302.0
105.80	300.0
106.90	131.0
115.70	257.0
116.80	468.0
117.80	281.0
118.80	457.0
127.80	288.0
128.90	118.0
129.80	262.0
134.80	125.0
136.80	144.0
140.90	677.0
142.80	767.0
144.80	121.0
145.70	155.0
147.90	192.0
154.90	194.0
156.90	169.0
171.80	656.0
172.90	607.0
173.80	76688.0
174.90	5377.0
175.80	73952.0
176.90	4777.0
177.90	128.0

Instrument :
MSVOA_V
ClientSampleId :
BFB63

m/z	Abundance
35.90	496.0
36.90	2827.0
37.90	2659.0
38.90	1178.0
39.90	212.0
43.90	1058.0
44.90	498.0
46.90	986.0
47.80	418.0
48.90	2766.0
49.90	13070.0
50.90	4180.0
52.00	194.0
54.90	232.0
55.90	894.0
56.90	1777.0
59.90	649.0
60.90	3398.0
61.90	3473.0
62.90	2743.0
63.90	248.0
64.90	555.0
66.90	174.0
67.90	7541.0
68.90	7842.0
69.90	568.0
71.90	411.0
72.90	3305.0
73.90	12570.0
74.90	39536.0
75.90	3521.0
76.90	629.0
77.80	408.0
78.80	1841.0
79.90	525.0
80.80	1959.0
81.80	477.0
85.90	141.0
86.90	4081.0
87.90	3492.0
90.80	230.0
91.90	2406.0
92.90	3139.0
93.90	9213.0
94.90	86768.0
96.00	5678.0
96.90	143.0
103.80	317.0
105.80	302.0
106.80	130.0
115.70	251.0
116.80	439.0
117.90	270.0
118.80	416.0
127.80	291.0
129.80	296.0
134.80	107.0
136.80	157.0
140.90	693.0
142.80	684.0
144.80	118.0
145.70	161.0
147.90	168.0
154.90	196.0
156.80	174.0
171.80	625.0
172.90	576.0
173.90	76008.0
174.90	5158.0
175.80	73448.0
176.90	4551.0
177.80	123.0

Instrument :
MSVOA_V
ClientSampleId :
BFB63