

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071023\
 Data File : VV031667.D
 Acq On : 10 Jul 2023 15:17
 Operator : SY/MD
 Sample : VIBLK223
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK223

Quant Time: Jul 11 01:09:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 07 02:34:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	146176	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	149867	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	78091	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	55721	57.276	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	114.560%	
7) Chloroethane-d5	1.532	69	46769	55.097	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	110.200%	
11) 1,1-Dichloroethene-d2	2.063	65	26122	52.099	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	104.200%	
21) 2-Butanone-d5	3.796	46	82288	104.460	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	104.460%	
24) Chloroform-d	4.256	84	119726	49.140	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.280%	
26) 1,2-Dichloroethane-d4	4.950	65	82570	54.166	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	108.340%	
32) Benzene-d6	4.966	84	202074	47.903	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	5.995	67	72704	48.627	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	97.260%	
41) Toluene-d8	7.249	98	163904	42.341	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	84.680%	
43) trans-1,3-Dichloroprop...	7.558	79	32209	44.190	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	88.380%	
47) 2-Hexanone-d5	8.030	63	48452	82.076	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	82.080%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	100576	44.566	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	89.140%	
66) 1,2-Dichlorobenzene-d4	11.567	152	67165	45.121	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.240%	
Target Compounds						
72) 1,2,3-Trichlorobenzene	13.686	180	2024	1.168	ug/L	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071023\
Data File : VV031667.D
Acq On : 10 Jul 2023 15:17
Operator : SY/MD
Sample : VIBLK223
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK223

Quant Time: Jul 11 01:09:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 07 02:34:55 2023
Response via : Initial Calibration

