

Data File : VV011856.D  
 Acq On : 11 Jul 2019 15:49  
 Operator : SY/MD  
 Sample : VSTDCCC0.5  
 Misc : 25ML/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTD0.522

Quant Time: Jul 12 03:43:45 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR070819SIM.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Thu Jul 11 05:02:52 2019  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	8095	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	7356	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3472	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	3864	0.44	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	88.00%
4) 1,2-Dichloroethane-d4	5.07	65	3032	0.46	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.00%
7) 1,2-Dichloropropane-d6	6.12	67	3376	0.47	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	94.00%
8) Toluene-d8	7.35	98	9243	0.50	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
10) 1,1,2,2-Tetrachloroethane-	10.25	84	1692	0.38	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	76.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.32	62	4385	0.522	ug/L 100
6) Trichloroethene	5.95	95	3704	0.547	ug/L 95
9) 1,2-Dibromoethane	8.40	107	2048	0.558	ug/L 100
12) 1,2,3-Trichloropropane	10.30	75	1947	0.515	ug/L 100
13) 1,2-Dibromo-3-chloropropan	12.46	75	299	0.431	ug/L 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011856.D  
Acq On : 11 Jul 2019 15:49  
Operator : SY/MD  
Sample : VSTDCCC0.5  
Misc : 25ML/MSVOA\_V/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampled :  
VSTD0.522

Quant Time: Jul 12 03:43:45 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR070819SIM.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Thu Jul 11 05:02:52 2019  
Response via : Initial Calibration

