

Data File : VV011857.D
Acq On : 11 Jul 2019 16:27
Operator : SY/MD
Sample : VV0711WBL01
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK10

Quant Time: Jul 12 03:48:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 12 03:44:53 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	7655	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	7145	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3372	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	3302	0.40	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	80.00%
4) 1,2-Dichloroethane-d4	5.07	65	3142	0.50	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
7) 1,2-Dichloropropane-d6	6.12	67	3532	0.51	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	102.00%
8) Toluene-d8	7.35	98	9487	0.53	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.00%
10) 1,1,2,2-Tetrachloroethane-	10.25	84	1749	0.40	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	80.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011857.D
Acq On : 11 Jul 2019 16:27
Operator : SY/MD
Sample : VV0711WBL01
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK10

Quant Time: Jul 12 03:48:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 12 03:44:53 2019
Response via : Initial Calibration

