

Data File : VV011860.D
Acq On : 11 Jul 2019 18:44
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.523

Quant Time: Jul 12 03:47:42 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 12 03:44:53 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	7138	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	6512	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3097	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	4344	0.56	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	112.00%
4) 1,2-Dichloroethane-d4	5.09	65	3055	0.52	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.00%
7) 1,2-Dichloropropane-d6	6.12	67	3347	0.53	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	106.00%
8) Toluene-d8	7.35	98	7995	0.49	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.00%
10) 1,1,2,2-Tetrachloroethane-	10.25	84	1726	0.44	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	88.00%

Target Compounds

						Qvalue
3) Vinyl chloride	1.32	62	4755	0.642	ug/L	99
6) Trichloroethene	5.95	95	3178	0.530	ug/L	98
9) 1,2-Dibromoethane	8.40	107	1792	0.552	ug/L	99
12) 1,2,3-Trichloropropane	10.30	75	1784	0.529	ug/L	99
13) 1,2-Dibromo-3-chloropropan	12.46	75	290	0.468	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011860.D
Acq On : 11 Jul 2019 18:44
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.523

Quant Time: Jul 12 03:47:42 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 12 03:44:53 2019
Response via : Initial Calibration

