

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071123\
 Data File : VV031679.D
 Acq On : 11 Jul 2023 15:20
 Operator : SY/MD
 Sample : VV0711MBL01
 Misc : 5.00g/5.0uL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK369

Quant Time: Jul 11 16:06:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 11 13:50:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	311009	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	324871	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	150000	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	130643	46.410	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 92.820%			
7) Chloroethane-d5	1.532	69	115136	56.520	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 113.040%			
11) 1,1-Dichloroethene-d2	2.060	65	64388	51.347	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 102.700%			
21) 2-Butanone-d5	3.793	46	175336	99.249	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 99.250%			
24) Chloroform-d	4.252	84	263213	46.852	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 93.700%			
26) 1,2-Dichloroethane-d4	4.944	65	166099	47.020	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 94.040%			
32) Benzene-d6	4.963	84	461509	47.380	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 94.760%			
36) 1,2-Dichloropropane-d6	5.992	67	164432	52.212	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 104.420%			
41) Toluene-d8	7.249	98	381665	43.580	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 87.160%			
43) trans-1,3-Dichloroprop...	7.558	79	68003	43.242	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 86.480%			
47) 2-Hexanone-d5	8.030	63	103940	88.506	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 88.510%			
56) 1,1,2,2-Tetrachloroeth...	10.159	84	217695	45.638	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 91.280%			
66) 1,2-Dichlorobenzene-d4	11.567	152	148328	53.912	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 107.820%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071123\
Data File : VV031679.D
Acq On : 11 Jul 2023 15:20
Operator : SY/MD
Sample : VV0711MBL01
Misc : 5.00g/5.0uL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK369

Quant Time: Jul 11 16:06:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071123WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 11 13:50:08 2023
Response via : Initial Calibration

