



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV071125\
Data File : VV038878.D
Acq On : 11 Jul 2025 07:51
Operator : SY/MD
Sample : BFB
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V071125W.M
Title : SW846 8260
Last Update : Mon Jul 14 09:44:44 2025



AutoFind: Scans 2786, 2787, 2788; Background Corrected with Scan 2775

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	12006	PASS
75	95	30	60	51.0	38445	PASS
95	95	100	100	100.0	75419	PASS
96	95	5	9	7.2	5394	PASS
173	174	0.00	2	0.9	590	PASS
174	95	50	100	86.4	65173	PASS
175	174	5	9	6.9	4527	PASS
176	174	95	101	95.0	61915	PASS
177	176	5	9	6.1	3754	PASS