

Data File : VV006593.D
Acq On : 12 Jul 2018 18:33
Operator : SY/MD
Sample : J3919-19
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF7

Quant Time: Jul 15 23:51:36 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 14 00:30:59 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	304463	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	278198	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	121326	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	64129	3.87	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.40%
7) Chloroethane-d5	1.58	69	63192	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.40%
11) 1,1-Dichloroethene-d2	2.13	63	89028	2.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.80%#
20) 2-Butanone-d5	3.95	46	194100	46.16	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.32%
24) Chloroform-d	4.41	84	146533	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.00%
26) 1,2-Dichloroethane-d4	5.09	65	70864	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.11	84	301569	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.12	67	97061	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.37	98	255068	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	7.67	79	27801	3.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.60%
46) 2-Hexanone-d5	8.14	63	177596	50.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.76%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	73848	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	107401	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds					Qvalue
13) Acetone	2.19	43	2820	1.060 ug/L	98
25) Chloroform	4.44	83	8244	0.230 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV006593.D
Acq On : 12 Jul 2018 18:33
Operator : SY/MD
Sample : J3919-19
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF7

Quant Time: Jul 15 23:51:36 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 14 00:30:59 2018
Response via : Initial Calibration

