

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071221\
 Data File : VV021618.D
 Acq On : 12 Jul 2021 11:48
 Operator : SY/MD
 Sample : VV0712WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK212

Quant Time: Jul 13 01:15:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 13 01:14:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	477556	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	441984	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	222504	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	165064	54.630	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 109.260%			
7) Chloroethane-d5	1.558	69	125590	50.857	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 101.720%			
11) 1,1-Dichloroethene-d2	2.104	63	210441	38.067	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 76.140%			
21) 2-Butanone-d5	3.866	46	246554	95.545	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 95.540%			
24) Chloroform-d	4.346	84	303191	47.973	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 95.940%			
26) 1,2-Dichloroethane-d4	5.027	65	188593	50.643	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 101.280%			
32) Benzene-d6	5.047	84	642206	51.750	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 103.500%			
36) 1,2-Dichloropropane-d6	6.069	67	199832	51.054	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 102.100%			
41) Toluene-d8	7.316	98	594438	51.794	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 103.580%			
43) trans-1,3-Dichloroprop...	7.622	79	100943	51.844	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 103.680%			
47) 2-Hexanone-d5	8.085	63	174603	96.169	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 96.170%			
56) 1,1,2,2-Tetrachloroeth...	10.217	84	269246	51.056	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 102.120%			
66) 1,2-Dichlorobenzene-d4	11.625	152	232182	52.341	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 104.680%			
Target Compounds						Qvalue
16) Methylene chloride	2.506	84	3836	1.036	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071221\
Data File : VV021618.D
Acq On : 12 Jul 2021 11:48
Operator : SY/MD
Sample : VV0712WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK212

Quant Time: Jul 13 01:15:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 13 01:14:23 2021
Response via : Initial Calibration

