

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071221\
 Data File : VV021625.D
 Acq On : 12 Jul 2021 14:49
 Operator : SY/MD
 Sample : M3008-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGE66

Quant Time: Jul 13 01:17:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 13 01:14:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	491056	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	450753	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	219296	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	155221	49.960	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.920%
7) Chloroethane-d5	1.558	69	118192	46.545	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.100%
11) 1,1-Dichloroethene-d2	2.105	63	198091	34.848	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.700%
21) 2-Butanone-d5	3.873	46	245851	92.653	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.650%
24) Chloroform-d	4.349	84	290873	44.759	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.520%
26) 1,2-Dichloroethane-d4	5.034	65	179982	47.002	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.000%
32) Benzene-d6	5.050	84	603641	47.696	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.400%
36) 1,2-Dichloropropane-d6	6.072	67	189304	47.424	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.840%
41) Toluene-d8	7.317	98	555284	47.441	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.880%
43) trans-1,3-Dichloroprop...	7.625	79	93921	47.299	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.600%
47) 2-Hexanone-d5	8.088	63	170293	91.970	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.970%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	260138	48.369	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.740%
66) 1,2-Dichlorobenzene-d4	11.625	152	213114	48.745	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.500%
Target Compounds						
					Qvalue	
13) Acetone	2.159	43	6837	3.117	ug/L	# 100
25) Chloroform	4.378	83	182554	28.731	ug/L	99
38) Bromodichloromethane	6.513	83	33011	6.851	ug/L	96
49) Dibromochloromethane	8.252	129	10394	2.577	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071221\
Data File : VV021625.D
Acq On : 12 Jul 2021 14:49
Operator : SY/MD
Sample : M3008-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGE66

Quant Time: Jul 13 01:17:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 13 01:14:23 2021
Response via : Initial Calibration

