

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071223\
 Data File : VV031687.D
 Acq On : 12 Jul 2023 11:50
 Operator : SY/MD
 Sample : VV0712WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK372

Quant Time: Jul 13 04:58:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 13 04:58:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	281227	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	300640	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	132993	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	116988	45.960	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.920%
7) Chloroethane-d5	1.532	69	105590	57.324	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	114.640%
11) 1,1-Dichloroethene-d2	2.063	65	59011	52.042	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.080%
21) 2-Butanone-d5	3.796	46	162370	99.819	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.820%
24) Chloroform-d	4.256	84	253202	49.843	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.680%
26) 1,2-Dichloroethane-d4	4.947	65	160207	50.155	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.320%
32) Benzene-d6	4.966	84	435766	48.343	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.680%
36) 1,2-Dichloropropane-d6	5.995	67	158396	54.349	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.700%
41) Toluene-d8	7.249	98	354288	43.715	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.420%
43) trans-1,3-Dichloroprop...	7.558	79	64112	44.054	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.100%
47) 2-Hexanone-d5	8.030	63	97443	87.229	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.230%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	208775	47.296	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	11.567	152	137986	56.567	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071223\
Data File : VV031687.D
Acq On : 12 Jul 2023 11:50
Operator : SY/MD
Sample : VV0712WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK372

Quant Time: Jul 13 04:58:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071123WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 13 04:58:07 2023
Response via : Initial Calibration

