

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071318\
 Data File : VV006601.D
 Acq On : 13 Jul 2018 19:02
 Operator : SY/MD
 Sample : J3919-17RE
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF5RE

Quant Time: Jul 14 01:57:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 14 00:30:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	337444	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	308568	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	133754	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	89817	4.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.00%
7) Chloroethane-d5	1.58	69	72928	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.00%
11) 1,1-Dichloroethene-d2	2.13	63	118414	3.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.40%
20) 2-Butanone-d5	3.97	46	192355	41.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.56%
24) Chloroform-d	4.41	84	152638	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
26) 1,2-Dichloroethane-d4	5.09	65	74208	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.60%
32) Benzene-d6	5.10	84	331029	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
36) 1,2-Dichloropropane-d6	6.12	67	101287	4.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.80%
41) Toluene-d8	7.37	98	295447	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
43) trans-1,3-Dichloropropene-	7.67	79	30830	3.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.60%
46) 2-Hexanone-d5	8.14	63	161908	41.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.82%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	71214	3.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	107780	4.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.60%

Target Compounds					Ovalue
13) Acetone	2.22	43	10303	3.50 ug/L	96
25) Chloroform	4.43	83	29864	0.75 ug/L	98
34) Trichloroethene	5.96	95	41560	1.76 ug/L	98
47) Tetrachloroethene	8.02	164	7172	0.37 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071318\
Data File : VV006601.D
Acq On : 13 Jul 2018 19:02
Operator : SY/MD
Sample : J3919-17RE
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF5RE

Quant Time: Jul 14 01:57:23 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 14 00:30:59 2018
Response via : Initial Calibration

