

Data File : VV006598.D
Acq On : 13 Jul 2018 17:13
Operator : SY/MD
Sample : VV0713WBL01
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK64

Quant Time: Jul 14 00:33:48 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 14 00:30:59 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	351145	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	314473	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	138869	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	96296	5.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.80%
7) Chloroethane-d5	1.58	69	78018	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.60%
11) 1,1-Dichloroethene-d2	2.13	63	126471	3.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.20%
20) 2-Butanone-d5	3.98	46	178791	36.87	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	73.74%
24) Chloroform-d	4.40	84	155125	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.60%
26) 1,2-Dichloroethane-d4	5.09	65	73931	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.80%
32) Benzene-d6	5.10	84	351453	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.12	67	103267	4.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.80%
41) Toluene-d8	7.36	98	314594	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
43) trans-1,3-Dichloropropene-	7.67	79	31714	4.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.40%
46) 2-Hexanone-d5	8.14	63	149896	37.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.22%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	67172	3.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	111139	4.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV006598.D
Acq On : 13 Jul 2018 17:13
Operator : SY/MD
Sample : VV0713WBL01
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK64

Quant Time: Jul 14 00:33:48 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 14 00:30:59 2018
Response via : Initial Calibration

