

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW071321\
 Data File : VW021635.D
 Acq On : 13 Jul 2021 10:57
 Operator : SY/MD
 Sample : M2970-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK002

Quant Time: Jul 14 03:33:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 14 03:32:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	440998	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	409819	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	200501	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	137741	49.366	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.740%
7) Chloroethane-d5	1.558	69	110520	48.464	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.920%
11) 1,1-Dichloroethene-d2	2.105	63	181181	35.491	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.980%
21) 2-Butanone-d5	3.873	46	249939	104.886	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.890%
24) Chloroform-d	4.349	84	269191	46.125	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.240%
26) 1,2-Dichloroethane-d4	5.034	65	175793	51.119	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.240%
32) Benzene-d6	5.050	84	569816	49.520	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.040%
36) 1,2-Dichloropropane-d6	6.072	67	182320	50.236	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.480%
41) Toluene-d8	7.317	98	522099	49.062	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.120%
43) trans-1,3-Dichloroprop...	7.622	79	89397	49.518	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.040%
47) 2-Hexanone-d5	8.088	63	168491	100.086	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.090%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	256026	52.360	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.720%
66) 1,2-Dichlorobenzene-d4	11.625	152	199761	49.974	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.940%
Target Compounds						
25) Chloroform	4.381	83	4422	0.775	ug/L	Qvalue # 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071321\
Data File : VV021635.D
Acq On : 13 Jul 2021 10:57
Operator : SY/MD
Sample : M2970-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK002

Quant Time: Jul 14 03:33:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 14 03:32:27 2021
Response via : Initial Calibration

