

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071321\
 Data File : VV021643.D
 Acq On : 13 Jul 2021 14:08
 Operator : SY/MD
 Sample : M3008-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jul 14 03:34:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 14 03:32:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	441881	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	414710	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	202088	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	135896	48.608	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.220%
7) Chloroethane-d5	1.561	69	107165	46.899	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.800%
11) 1,1-Dichloroethene-d2	2.105	63	178523	34.900	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.800%
21) 2-Butanone-d5	3.873	46	247723	103.748	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.750%
24) Chloroform-d	4.349	84	270093	46.187	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.380%
26) 1,2-Dichloroethane-d4	5.034	65	174850	50.743	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.480%
32) Benzene-d6	5.050	84	568860	48.854	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.700%
36) 1,2-Dichloropropane-d6	6.072	67	184132	50.137	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.280%
41) Toluene-d8	7.317	98	523037	48.570	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.140%
43) trans-1,3-Dichloroprop...	7.622	79	88744	48.576	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.160%
47) 2-Hexanone-d5	8.088	63	164354	96.477	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.480%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	252565	51.043	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.080%
66) 1,2-Dichlorobenzene-d4	11.625	152	201447	50.000	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.000%
Target Compounds						
25) Chloroform	4.378	83	4402	0.770	ug/L	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071321\
Data File : VV021643.D
Acq On : 13 Jul 2021 14:08
Operator : SY/MD
Sample : M3008-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jul 14 03:34:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 14 03:32:27 2021
Response via : Initial Calibration

