

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071420\
 Data File : VV017481.D
 Acq On : 14 Jul 2020 14:07
 Operator : SY/MD
 Sample : L3281-03MSD
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAXW3MSD

Quant Time: Jul 15 03:32:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 15 02:49:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	140173	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	137635	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	69563	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40457	4.96	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.20%
7) Chloroethane-d5	1.58	69	32930	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.60%
11) 1,1-Dichloroethene-d2	2.13	63	87827	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.60%
20) 2-Butanone-d5	3.93	46	87352	52.21	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.42%
24) Chloroform-d	4.38	84	84171	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.06	65	48584	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.08	84	154194	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
36) 1,2-Dichloropropane-d6	6.10	67	42930	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.34	98	145873	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
45) trans-1,3-Dichloropropene-	7.64	79	19731	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
48) 2-Hexanone-d5	8.11	63	63720	49.86	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.72%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	29051	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	59573	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	42151	4.998 ug/L	82
13) Acetone	2.21	43	7471	6.151 ug/L	74
33) Benzene	5.13	78	185419	5.063 ug/L	100
34) Trichloroethene	5.94	95	52621	4.706 ug/L	97
42) Toluene	7.41	91	216224	5.347 ug/L	98
53) Chlorobenzene	8.90	112	138858	5.126 ug/L	97
54) Ethylbenzene	9.04	91	4502	0.097 ug/L	84
55) m,p-xylene	9.16	106	1251	0.071 ug/L #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV071420\
Data File : VV017481.D
Acq On : 14 Jul 2020 14:07
Operator : SY/MD
Sample : L3281-03MSD
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAXW3MSD

Quant Time: Jul 15 03:32:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 15 02:49:58 2020
Response via : Initial Calibration

