

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071421\
 Data File : VV021651.D
 Acq On : 14 Jul 2021 14:24
 Operator : SY/MD
 Sample : M3033-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG7M0

Quant Time: Jul 15 01:59:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 15 01:57:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	431532	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	397154	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	197769	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	134896	49.407	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	98.820%	
7) Chloroethane-d5	1.561	69	109744	49.180	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	98.360%	
11) 1,1-Dichloroethene-d2	2.105	63	181276	36.288	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	72.580%	
21) 2-Butanone-d5	3.873	46	239966	102.910	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	102.910%	
24) Chloroform-d	4.349	84	276474	48.412	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.820%	
26) 1,2-Dichloroethane-d4	5.034	65	174519	51.862	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	103.720%	
32) Benzene-d6	5.050	84	578326	51.863	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	103.720%	
36) 1,2-Dichloropropane-d6	6.072	67	184139	52.355	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	104.720%	
41) Toluene-d8	7.317	98	533130	51.696	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	103.400%	
43) trans-1,3-Dichloroprop...	7.622	79	88370	50.510	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	101.020%	
47) 2-Hexanone-d5	8.088	63	155098	95.068	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	95.070%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	241028	50.864	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	101.720%	
66) 1,2-Dichlorobenzene-d4	11.625	152	211528	53.649	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	107.300%	
Target Compounds						
					Qvalue	
13) Acetone	2.159	43	18268	9.476	ug/L #	100
48) 2-Hexanone	8.146	43	15846	4.347	ug/L #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071421\
Data File : VV021651.D
Acq On : 14 Jul 2021 14:24
Operator : SY/MD
Sample : M3033-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG7M0

Quant Time: Jul 15 01:59:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062921WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 15 01:57:12 2021
Response via : Initial Calibration

