

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071423\
 Data File : VV031703.D
 Acq On : 14 Jul 2023 10:09
 Operator : SY/MD
 Sample : VV0714WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK349

Quant Time: Jul 15 05:26:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jul 15 05:25:11 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.542	114	163087	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.792	117	178931	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.191	152	72918	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	76979	5.330	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.600%	
7) Chloroethane-d5		1.532	69	74192	5.667	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.400%	
11) 1,1-Dichloroethene-d2		2.060	65	39907	5.384	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	107.600%	
20) 2-Butanone-d5		3.818	46	187672	50.862	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.720%	
24) Chloroform-d		4.256	84	163190	5.304	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.000%	
26) 1,2-Dichloroethane-d4		4.950	65	84715	5.391	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.800%	
32) Benzene-d6		4.966	84	258371	5.058	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.200%	
36) 1,2-Dichloropropane-d6		5.995	67	99626	5.262	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%	
41) Toluene-d8		7.249	98	187297	3.992	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%	
43) trans-1,3-Dichloroprop...		7.564	79	28272	4.728	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.600%	
46) 2-Hexanone-d5		8.034	63	124137	44.700	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.400%	
56) 1,1,2,2-Tetrachloroeth...		10.159	84	71839	5.172	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.400%	
66) 1,2-Dichlorobenzene-d4		11.567	152	70715	5.569	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.400%	
Target Compounds							Qvalue
16) Methylene chloride		2.449	84	4628	0.274	ug/L	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071423\
Data File : VV031703.D
Acq On : 14 Jul 2023 10:09
Operator : SY/MD
Sample : VV0714WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK349

Quant Time: Jul 15 05:26:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jul 15 05:25:11 2023
Response via : Initial Calibration

