

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071519\
 Data File : VV011884.D
 Acq On : 15 Jul 2019 12:34
 Operator : SY/MD
 Sample : VV0715WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK48

Quant Time: Jul 16 04:04:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 16 03:56:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	181053	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	177725	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	76629	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	59005	4.34	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.80%
7) Chloroethane-d5	1.58	69	50643	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.13	63	86851	4.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.00%
20) 2-Butanone-d5	3.96	46	144308	50.54	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.08%
24) Chloroform-d	4.40	84	112377	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
26) 1,2-Dichloroethane-d4	5.08	65	61547	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
32) Benzene-d6	5.10	84	227549	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
36) 1,2-Dichloropropane-d6	6.12	67	71030	4.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.20%
41) Toluene-d8	7.36	98	196616	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
43) trans-1,3-Dichloropropene-	7.67	79	22469	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	8.14	63	93524	43.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.18%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	40395	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	68268	4.82	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV071519\
Data File : VV011884.D
Acq On : 15 Jul 2019 12:34
Operator : SY/MD
Sample : VV0715WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK48

Quant Time: Jul 16 04:04:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 16 03:56:26 2019
Response via : Initial Calibration

