

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071520\
 Data File : VV071500.D
 Acq On : 15 Jul 2020 10:16
 Operator : SY/MD
 Sample : VV0715WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK34

Quant Time: Jul 16 01:57:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 16 01:29:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	151667	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	143535	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	72171	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	37346	4.23	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.60%
7) Chloroethane-d5	1.58	69	31871	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.80%
11) 1,1-Dichloroethene-d2	2.12	63	60177	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.40%
20) 2-Butanone-d5	3.97	46	81409	44.97	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.94%
24) Chloroform-d	4.38	84	84357	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.20%
26) 1,2-Dichloroethane-d4	5.06	65	48783	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
32) Benzene-d6	5.07	84	151921	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.10	67	42634	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.80%
41) Toluene-d8	7.34	98	139209	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
45) trans-1,3-Dichloropropene-	7.65	79	19116	4.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.60%
48) 2-Hexanone-d5	8.12	63	55709	41.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.60%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	28537	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	57103	4.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071520\
Data File : VV017500.D
Acq On : 15 Jul 2020 10:16
Operator : SY/MD
Sample : VV0715WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK34

Quant Time: Jul 16 01:57:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 16 01:29:31 2020
Response via : Initial Calibration

