

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV071525\
 Data File : VV038892.D
 Acq On : 15 Jul 2025 10:07
 Operator : SY/MD
 Sample : VV0715WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0715WBL02

Quant Time: Jul 16 01:42:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V071125W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 16 01:18:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	498048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	865959	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	775960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	336351	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.940	65	340505	55.955	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	111.920%
35) Dibromofluoromethane	4.503	113	285446	48.347	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	96.700%
50) Toluene-d8	7.236	98	924379	51.796	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	103.600%
62) 4-Bromofluorobenzene	10.001	95	336144	46.757	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	93.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV071525\
Data File : VV038892.D
Acq On : 15 Jul 2025 10:07
Operator : SY/MD
Sample : VV0715WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0715WBL02

Quant Time: Jul 16 01:42:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V071125W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 16 01:18:54 2025
Response via : Initial Calibration

