

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV071525\
 Data File : VV038890.D
 Acq On : 15 Jul 2025 08:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 16 03:05:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V071125W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 16 03:05:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.439	0.375	14.6	98	0.00
3 P	Chloromethane	0.474	0.377	20.5#	96	0.00
4 C	Vinyl Chloride	0.543	0.466	14.2#	104	0.00
5 T	Bromomethane	0.407	0.337	17.2	102	0.00
6 T	Chloroethane	0.363	0.329	9.4	106	0.00
7 T	Trichlorofluoromethane	0.961	0.881	8.3	103	0.00
8 T	Diethyl Ether	0.345	0.337	2.3	109	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.574	0.548	4.5	105	0.00
10 T	Methyl Iodide	0.660	0.650	1.5	101	0.00
11 T	Tert butyl alcohol	0.141	0.130	7.8	99	0.00
12 CM	1,1-Dichloroethene	0.523	0.461	11.9#	102	0.00
13 T	Acrolein	0.020	0.020	0.0	100	0.00
14 T	Allyl chloride	0.834	0.787	5.6	107	0.00
15 T	Acrylonitrile	0.360	0.349	3.1	107	0.00
16 T	Acetone	0.264	0.268	-1.5	105	0.00
17 T	Carbon Disulfide	1.092	0.770	29.5#	94	0.00
18 T	Methyl Acetate	0.894	0.978	-9.4	109	0.00
19 T	Methyl tert-butyl Ether	1.963	1.924	2.0	106	0.00
20 T	Methylene Chloride	0.671	0.580	13.6	104	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.474	14.9	100	0.00
22 T	Diisopropyl ether	1.666	1.746	-4.8	107	0.00
23 T	Vinyl Acetate	1.307	1.363	-4.3	107	0.00
24 P	1,1-Dichloroethane	1.091	1.033	5.3	104	0.00
25 T	2-Butanone	0.379	0.396	-4.5	105	0.00
26 T	2,2-Dichloropropane	0.955	0.938	1.8	105	0.00
27 T	cis-1,2-Dichloroethene	0.640	0.626	2.2	106	0.00
28 T	Bromochloromethane	0.422	0.437	-3.6	105	0.00
29 T	Tetrahydrofuran	0.250	0.257	-2.8	105	0.00
30 C	Chloroform	1.203	1.128	6.2#	103	0.00
31 T	Cyclohexane	0.716	0.632	11.7	102	0.00
32 T	1,1,1-Trichloroethane	1.050	0.979	6.8	104	0.00
33 S	1,2-Dichloroethane-d4	0.611	0.514	15.9	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.341	0.305	10.6	97	0.00
36 T	1,1-Dichloropropene	0.392	0.385	1.8	101	0.00
37 T	Ethyl Acetate	0.461	0.474	-2.8	103	0.00
38 T	Carbon Tetrachloride	0.520	0.505	2.9	103	0.00
39 T	Methylcyclohexane	0.462	0.431	6.7	99	0.00
40 TM	Benzene	1.328	1.304	1.8	102	0.00
41 T	Methacrylonitrile	0.212	0.240	-13.2	106	0.00
42 TM	1,2-Dichloroethane	0.506	0.495	2.2	104	0.00
43 T	Isopropyl Acetate	0.730	0.755	-3.4	105	0.00
44 TM	Trichloroethene	0.347	0.326	6.1	103	0.00
45 C	1,2-Dichloropropane	0.351	0.359	-2.3#	106	0.00
46 T	Dibromomethane	0.271	0.266	1.8	103	0.00
47 T	Bromodichloromethane	0.576	0.584	-1.4	104	0.00
48 T	Methyl methacrylate	0.317	0.359	-13.2	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV071525\
 Data File : VV038890.D
 Acq On : 15 Jul 2025 08:54
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 16 03:05:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V071125W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 16 03:05:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	83	0.00
50 S	Toluene-d8	1.030	0.900	12.6	93	0.00
51 T	4-Methyl-2-Pentanone	0.497	0.552	-11.1	106	0.00
52 CM	Toluene	0.837	0.873	-4.3#	104	0.00
53 T	t-1,3-Dichloropropene	0.521	0.559	-7.3	106	0.00
54 T	cis-1,3-Dichloropropene	0.563	0.590	-4.8	106	0.00
55 T	1,1,2-Trichloroethane	0.400	0.397	0.8	103	0.00
56 T	Ethyl methacrylate	0.475	0.556	-17.1	105	0.00
57 T	1,3-Dichloropropane	0.600	0.626	-4.3	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.287	0.366	-27.5#	108	0.00
59 T	2-Hexanone	0.370	0.403	-8.9	104	0.00
60 T	Dibromochloromethane	0.471	0.488	-3.6	107	0.00
61 T	1,2-Dibromoethane	0.403	0.405	-0.5	104	0.00
62 S	4-Bromofluorobenzene	0.415	0.407	1.9	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.328	0.306	6.7	102	0.00
65 PM	Chlorobenzene	1.093	1.093	0.0	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.436	0.435	0.2	104	0.00
67 C	Ethyl Benzene	1.622	1.806	-11.3#	104	0.00
68 T	m/p-Xylenes	0.632	0.726	-14.9	103	0.00
69 T	o-Xylene	0.596	0.683	-14.6	105	0.00
70 T	Styrene	1.054	1.273	-20.8#	106	0.00
71 P	Bromoform	0.354	0.380	-7.3	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	2.988	3.269	-9.4	104	0.00
74 T	N-amyl acetate	1.094	1.237	-13.1	106	0.00
75 P	1,1,2,2-Tetrachloroethane	1.251	1.213	3.0	104	0.00
76 T	1,2,3-Trichloropropane	0.930	0.930	0.0	107	0.00
77 T	Bromobenzene	0.784	0.784	0.0	102	0.00
78 T	n-propylbenzene	3.492	3.912	-12.0	106	0.00
79 T	2-Chlorotoluene	2.221	2.326	-4.7	104	0.00
80 T	1,3,5-Trimethylbenzene	2.519	2.809	-11.5	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.381	0.416	-9.2	110	0.00
82 T	4-Chlorotoluene	2.482	2.759	-11.2	105	0.00
83 T	tert-Butylbenzene	2.641	2.889	-9.4	105	0.00
84 T	1,2,4-Trimethylbenzene	2.479	2.834	-14.3	104	0.00
85 T	sec-Butylbenzene	3.371	3.794	-12.5	106	0.00
86 T	p-Isopropyltoluene	2.862	3.204	-11.9	105	0.00
87 T	1,3-Dichlorobenzene	1.600	1.627	-1.7	105	0.00
88 T	1,4-Dichlorobenzene	1.740	1.685	3.2	106	0.00
89 T	n-Butylbenzene	2.613	3.055	-16.9	107	0.00
90 T	Hexachloroethane	0.681	0.651	4.4	106	0.00
91 T	1,2-Dichlorobenzene	1.553	1.579	-1.7	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.283	0.266	6.0	98	0.00
93 T	1,2,4-Trichlorobenzene	0.917	1.053	-14.8	108	0.00
94 T	Hexachlorobutadiene	0.472	0.481	-1.9	107	0.00
95 T	Naphthalene	3.012	3.680	-22.2#	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV071525\
Data File : VV038890.D
Acq On : 15 Jul 2025 08:54
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
LabSampleId :
VSTDCCC050

Quant Time: Jul 16 03:05:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V071125W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 16 03:05:29 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.950	1.090	-14.7	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6