

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV071625\
 Data File : VV038902.D
 Acq On : 16 Jul 2025 11:49
 Operator : SY/MD
 Sample : VV0716WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0716WBL01

Quant Time: Jul 17 06:12:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V071125W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 06:08:37 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.597	168	528643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	928044	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	804881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	299785	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.937	65	368229	57.009	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	114.020%
35) Dibromofluoromethane	4.500	113	309285	48.880	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	97.760%
50) Toluene-d8	7.236	98	990147	51.770	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	103.540%
62) 4-Bromofluorobenzene	10.001	95	331749	43.058	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	86.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV071625\
Data File : VV038902.D
Acq On : 16 Jul 2025 11:49
Operator : SY/MD
Sample : VV0716WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0716WBL01

Quant Time: Jul 17 06:12:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V071125W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 06:08:37 2025
Response via : Initial Calibration

