

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071719\
 Data File : VV011908.D
 Acq On : 17 Jul 2019 13:37
 Operator : SY/MD
 Sample : VV0717WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK50

Quant Time: Jul 18 02:53:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 18 02:17:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	174521	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	168383	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	74189	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	59213	4.62	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.40%
7) Chloroethane-d5	1.58	69	48345	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.00%
11) 1,1-Dichloroethene-d2	2.13	63	80212	3.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.40%
20) 2-Butanone-d5	3.96	46	155013	47.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.76%
24) Chloroform-d	4.40	84	113847	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
26) 1,2-Dichloroethane-d4	5.08	65	62940	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.10	84	229208	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
36) 1,2-Dichloropropane-d6	6.12	67	70846	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.36	98	201776	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	7.67	79	23315	4.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.00%
46) 2-Hexanone-d5	8.14	63	95924	42.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.80%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	40015	4.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	71754	4.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV071719\
Data File : VV011908.D
Acq On : 17 Jul 2019 13:37
Operator : SY/MD
Sample : VV0717WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK50

Quant Time: Jul 18 02:53:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 18 02:17:52 2019
Response via : Initial Calibration

