

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071818\
 Data File : VV006634.D
 Acq On : 18 Jul 2018 17:36
 Operator : SY/MD
 Sample : J3983-10DL 4X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB180DL

Quant Time: Jul 19 03:09:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 19 02:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	249036	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	229516	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	105668	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	96868	4.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.00%
7) Chloroethane-d5	1.58	69	82850	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.13	63	127074	3.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.20%
20) 2-Butanone-d5	3.97	46	215586	64.66	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	129.32%
24) Chloroform-d	4.41	84	164040	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.00%
26) 1,2-Dichloroethane-d4	5.09	65	79952	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.60%
32) Benzene-d6	5.10	84	348148	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
36) 1,2-Dichloropropane-d6	6.12	67	110442	4.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.00%
41) Toluene-d8	7.36	98	303312	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.80%
43) trans-1,3-Dichloropropene-	7.67	79	35070	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
46) 2-Hexanone-d5	8.14	63	189929	41.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.76%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	78280	4.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	111251	4.44	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.80%

Target Compounds						Ovalue
5) Vinyl chloride	1.32	62	213790	7.14	ug/L	# 44
18) trans-1,2-Dichloroethene	2.79	96	88658	4.04	ug/L	98
22) cis-1,2-Dichloroethene	3.97	96	10114	0.44	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV071818\
Data File : VV006634.D
Acq On : 18 Jul 2018 17:36
Operator : SY/MD
Sample : J3983-10DL 4X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB180DL

Quant Time: Jul 19 03:09:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 19 02:39:44 2018
Response via : Initial Calibration

