



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071822\
 Data File : VV026804.D
 Acq On : 18 Jul 2022 09:40
 Operator : SY/MD
 Sample : BFB347
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB347

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Tue Jul 19 03:06:10 2022



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	20931	PASS
75	95	30	80	49.1	62211	PASS
95	95	100	100	100.0	126739	PASS
96	95	5	9	6.4	8060	PASS
173	174	0.00	2	1.5	1471	PASS
174	95	50	120	79.1	100277	PASS
175	174	5	9	7.6	7659	PASS
176	174	95	105	95.0	95301	PASS
177	176	5	10	6.8	6485	PASS

m/z	Abundance
39.90	376.0
44.00	660.0

Instrument :
MSVOA_V
ClientSampleId :
BFB347

m/z	Abundance
36.00	806.0
37.00	5031.0
38.00	3779.0
39.00	2044.0
39.80	502.0
40.80	103.0
44.00	1132.0
45.00	765.0
45.80	244.0
46.30	249.0
47.10	1532.0
48.00	638.0
49.00	4564.0
50.00	20440.0
51.00	6454.0
52.00	369.0
52.90	215.0
54.00	110.0
55.00	310.0
56.10	1441.0
57.00	3215.0
57.90	111.0
60.00	1112.0
61.00	5765.0
62.00	5342.0
63.00	5145.0
63.90	1010.0
65.00	1106.0
66.00	147.0
67.10	327.0
68.00	11393.0
69.00	11377.0
70.00	954.0
72.00	655.0
73.00	5078.0
74.00	19800.0
75.00	59856.0
76.00	5164.0
77.10	856.0
77.80	799.0
78.90	5240.0
80.00	1298.0
80.90	5375.0
81.90	1040.0
85.60	148.0
86.00	178.0
87.00	5069.0
87.90	5063.0
90.90	482.0
91.90	3500.0
93.00	5222.0
94.00	14312.0
95.00	122120.0
96.10	7700.0
97.00	224.0
103.90	713.0
105.00	178.0
105.90	569.0
111.10	124.0
115.00	177.0
116.00	495.0
117.00	1016.0
117.90	555.0
118.90	914.0
123.80	142.0
127.80	431.0
128.90	204.0
130.00	451.0
130.90	177.0
134.80	250.0
136.70	197.0
141.00	1059.0
141.90	177.0
142.90	2217.0
143.90	141.0
144.90	1385.0
145.90	196.0
147.00	116.0
147.70	286.0
150.00	127.0
154.90	343.0
156.90	230.0
159.00	146.0
160.90	158.0
171.00	320.0
171.90	1251.0
173.10	1416.0
174.00	93624.0
175.00	7390.0
176.00	88760.0
177.00	6291.0
177.90	142.0

Instrument :
MSVOA_V
ClientSampleId :
BFB347

m/z	Abundance
35.90	957.0
37.00	5243.0
38.00	4062.0
39.00	2013.0
39.90	530.0
40.90	117.0
44.00	1195.0
45.00	896.0
45.80	246.0
47.10	1536.0
48.00	631.0
49.00	4927.0
50.00	21560.0
51.00	6697.0
52.00	326.0
52.90	242.0
54.00	143.0
55.00	286.0
56.10	1577.0
57.00	3304.0
57.90	130.0
60.00	1109.0
61.00	6006.0
62.00	5623.0
63.00	5343.0
64.00	1180.0
65.00	1167.0
66.00	184.0
67.10	397.0
68.00	12067.0
69.00	12359.0
70.00	1035.0
72.00	640.0
73.00	5656.0
74.00	21344.0
75.00	64336.0
76.00	5603.0
77.10	980.0
77.80	839.0
78.90	5691.0
79.90	1517.0
80.90	5630.0
81.90	1055.0
85.60	188.0
86.00	198.0
87.00	5279.0
87.90	5422.0
90.80	478.0
92.00	3831.0
93.00	5732.0
94.00	15133.0
95.00	130944.0
96.10	8294.0
97.00	231.0
103.90	702.0
105.10	179.0
105.90	552.0
107.90	110.0
111.10	112.0
111.90	119.0
112.90	131.0
114.90	174.0
116.00	472.0
117.00	1093.0
117.90	622.0
118.90	950.0
123.80	196.0
127.80	490.0
129.00	206.0
129.80	478.0
130.90	182.0
134.80	280.0
136.90	197.0
141.00	1125.0
141.80	177.0
142.90	2520.0
143.90	160.0
144.90	1464.0
145.90	206.0
147.00	160.0
147.80	279.0
150.00	169.0
154.90	392.0
157.00	264.0
158.90	173.0
160.90	187.0
171.00	366.0
171.90	1482.0
173.10	1517.0
174.00	103512.0
175.00	7940.0
176.00	98504.0

Instrument :
MSVOA_V
ClientSampleId :
BFB347

177.006695.0

178.00141.0

Instrument :
MSVOA_V
ClientSampleId :
BFB347

m/z	Abundance
35.90	1030.0
37.00	4900.0
38.00	4135.0
39.00	1917.0
39.90	510.0
40.80	148.0
44.00	1102.0
45.00	933.0
45.90	254.0
47.10	1371.0
47.90	596.0
49.00	4805.0
50.00	20792.0
51.00	6219.0
52.00	281.0
52.90	236.0
54.00	145.0
55.00	268.0
56.10	1564.0
57.00	3004.0
57.90	132.0
60.00	988.0
61.00	5697.0
62.00	5538.0
63.00	5214.0
64.00	1279.0
65.00	1201.0
66.00	181.0
67.10	419.0
68.00	11494.0
69.00	12098.0
70.00	1111.0
71.90	633.0
73.00	5788.0
74.00	20608.0
75.00	62440.0
76.00	5358.0
77.10	887.0
77.90	682.0
78.90	5531.0
79.90	1521.0
80.90	5546.0
81.90	1032.0
82.80	116.0
85.70	212.0
87.00	4981.0
87.90	5342.0
90.80	420.0
92.00	3712.0
93.00	5773.0
94.00	14712.0
95.00	127152.0
96.10	8186.0
97.00	238.0
103.90	635.0
105.00	173.0
105.90	501.0
108.00	145.0
110.10	102.0
111.90	123.0
112.90	122.0
114.90	144.0
116.00	381.0
117.00	1094.0
117.90	646.0
118.90	921.0
123.80	185.0
127.90	512.0
129.00	209.0
129.80	463.0
130.80	169.0
134.80	286.0
136.90	207.0
141.00	1107.0
141.80	169.0
142.90	2403.0
143.90	170.0
144.90	1324.0
146.00	202.0
147.00	136.0
148.00	291.0
149.80	172.0
154.90	337.0
157.00	290.0
158.80	190.0
160.90	174.0
171.00	443.0
171.90	1548.0
173.00	1480.0
174.00	103696.0
175.00	7648.0
176.00	98640.0

Instrument :
MSVOA_V
ClientSampleId :
BFB347

177.006470.0

178.00169.0

Instrument :
MSVOA_V
ClientSampleId :
BFB347